

TISSUE PROCESSING UNIT – 1No.
Doc. No: ST/12/Tissue Processing/2016-17/25

Features:

- Audible alarms, error message and warning codes.
- Control panel with full protected keyboard and LCD.
- Separately programmable for each station.
- Instrument should have the option of interrupting an automatic process for reloading or removing cassettes if needed to before the end of a run.

Specifications:

Wax baths:

Number : 2
 Capacity : Approx. 2L
 Temperature range : +5 °C Ambient – 70°C (Excess temperature cutout)

Reagent containers:

Number : 10
 Capacity : Approx. 2L

Standard tissue basket:

Number : 1
 Capacity : 70 cassettes or above

Programs:

Number : 4 or more Programmable
 Programmable time per station : 99 h 59 min
 Delayed start : > 7 days
 Drain time : Approx.60s

Operating voltage : 230 v / 50HZ

- ❖ **Suitable Vacuum pump to be supplied.**
- ❖ **Branded suitable UPS with 1hr backup**
- ❖ **Warranty: 3 years comprehensive warranty and 2 years non-comprehensive warranty should provide at free of cost.**
- ❖ **Any other items required to make it a complete system to be quoted**